

Standard Name 221 of 221

↑

8B 1096-4 ARSC

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 08-Feb-20 Time: 22:37:15

Batch Name 2 of 2

↑

R1117235 B22020110AA-001 1-1 3050960/VII/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 19-Sep-21 Time: 09:24:11

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)


Measured by:

Delta

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
R1117235 B22020110AA-001 1-1 3050960/VII/21	D65 10 Deg	0.56	1.94	1.25	0.33	0.87	0.25	0.96		Warn	too light	too bright	too yellow
R1117235 B22020110AA-001 1-1 3050960/VII/21	F11 10 Deg	0.50	1.25	1.07	0.28	0.63	0.21	0.72	0.70	Pass	too light	too bright	too yellow
R1117235 B22020110AA-001 1-1 3050960/VII/21	A 10 Deg	0.88	1.74	1.87	0.46	0.89	0.40	1.08	0.65	Pass	too light	too bright	too yellow