

Standard Name 13 of 14

↑

21B 0160-4 FRD

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 22-Jun-21 Time: 10:56:23

Batch Name 1 of 2

↑

MTR17080 B21229601AA-003 1-1 3076083/X/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal


Date: 22-Nov-21 Time: 09:03:54

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
MTR17080 B21229601AA-003 1-1 3076083/X/21	D65 10 Deg	0.18	-0.53	0.17	0.09	-0.22	0.65	0.69		Pass	too light	too dull	too yellow
MTR17080 B21229601AA-003 1-1 3076083/X/21	F11 10 Deg	0.16	-0.50	0.16	0.08	-0.18	0.68	0.71	0.04	Pass	too light	too dull	too yellow
MTR17080 B21229601AA-003 1-1 3076083/X/21	A 10 Deg	0.12	-0.58	0.07	0.06	-0.25	0.57	0.62	0.13	Pass	too light	too dull	too yellow



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form