

Standard Name 1 of 36

↑ AE6N Ambient BlushF21 ↓

Visual Hue Request:

New Std Retrieve Std Store Std List Std

Instrument Setup: %R LAV SCI UV Cal
Date: 10-Aug-20 Time: 09:20:32


Batch Name 1 of 1

↑ 60024818 21016118 1-1 3022235/III/21 ↓

New Bat Retrieve Bat Store Bat List Bat

Instrument Setup: %R LAV SCI UV Cal
Date: 03-May-21 Time: 15:40:39

Batch Name
60024818 21016118 1-1 3022235/III/21
60024818 21016118 1-1 3022235/III/21



Single/All Zoom In Zoom Out CMC Plot Print Form

DataColor TOOLS 2.4.4 S/N NO. 8823242

Standard %R

400-450nm	36.25	32.21	27.39	23.68	23.29	20.31
460-510nm	22.94	30.28	33.06	32.54	31.80	29.77
520-570nm	27.86	28.76	30.35	30.35	30.41	35.22
580-630nm	44.59	54.51	61.38	64.96	64.68	64.09
640-690nm	64.17	65.50	70.89	77.82	84.67	88.29
700nm	91.63					

CMC DE Pass/Fail Tolerance: 0.80

I:c Ratio: (2.00 1.00)

III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments
D65 10 Deg	1.03	-0.09	-0.23	0.41	-0.12	-0.13	0.44		Pass	lighter	less saturated more red
F11 10 Deg	1.03	-0.71	-0.77	0.40	-0.50	-0.04	0.64	0.86	Warn	lighter	less saturated more red