

Standard Name 13 of 13

↑

9C 0293 CRD

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 23-May-19 Time: 13:48:16

Batch Name 3 of 3

↑

SJS7091 5484629 1-1 3078087/X/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV 400


Date: 28-Nov-21 Time: 13:45:31

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments	
SJS7091 5484629 1-1 3078087/X/21	D65 10 Deg	0.98	0.02	0.13	0.70	-0.19	0.02	0.73		Pass	too light	too dull	too red
SJS7091 5484629 1-1 3078087/X/21	F11 10 Deg	0.92	-0.53	-0.00	0.67	0.42	-0.58	0.98	0.54	Warn	too light	too bright	too green
SJS7091 5484629 1-1 3078087/X/21	A 10 Deg	1.01	0.46	0.18	0.73	0.32	0.60	0.99	0.40	Pass	too light	too bright	too red



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form