

Standard Name 33 of 42

↑

21B 0282-2 ARL

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 26-Aug-21 Time: 17:03:42

Batch Name 2 of 2

↑

R1117150 B22009610AA-001 4-1 3053190/VII/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 08-Sep-21 Time: 11:36:55

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)


Measured by:

Delta 

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
R1117150 B22009610AA-001 4-1 3053190/VII/21	D65 10 Deg	-0.32	0.13	-0.00	-0.13	-0.17	-0.07	0.22		Pass	too dark	too dull	too yellow
R1117150 B22009610AA-001 4-1 3053190/VII/21	F11 10 Deg	-0.33	-0.04	-0.02	-0.13	0.04	0.04	0.14	0.17	Pass	too dark	too bright	too blue
R1117150 B22009610AA-001 4-1 3053190/VII/21	A 10 Deg	-0.30	0.11	0.05	-0.12	0.00	-0.18	0.22	0.06	Pass	too dark	too bright	too yellow