

Standard Name 131 of 140

↑

17109/220949L-B

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 17-Feb-23 Time: 13:36:53

Batch Name 1 of 2

↑

3110237IR-01 84386 EXP2300604 LILAC ETHE

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 12-Sep-23 Time: 15:03:14


CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)

Measured by: 

▼

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments			
HJY18402 23110237IR-01 84386 EXP2300604 LILAC ETHE	D65 10 Deg	0.06	0.01	0.91	0.02	-0.54	0.60	0.81		Warn	too light	too dull	too red	
HJY18402 23110237IR-01 84386 EXP2300604 LILAC ETHE	F11 10 Deg	0.14	-0.05	1.05	0.06	-0.65	0.56	0.86	0.17	Warn	too light	too dull	too red	
HJY18402 23110237IR-01 84386 EXP2300604 LILAC ETHE	A 10 Deg	0.13	0.02	1.05	0.05	-0.69	0.65	0.95	0.16	Warn	too light	too dull	too red	



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form