

Standard Name 477 of 478

↑

19106/220831L-C

↓

Visual Hue Request:

| | | | | |
|---------|--------------|-----------|----------|------------|
| New Std | Retrieve Std | Store Std | List Std | Update Std |
|---------|--------------|-----------|----------|------------|

Instrument Setup:%R LAV SCI UV Cal

Date: 17-Jul-23 Time: 10:45:34

Batch Name 1 of 3

↑

30120130IR-01 83697 EXP23006050 MOJAVE T

↓

| | | | | |
|---------|--------------|-----------|----------|------------|
| New Bat | Retrieve Bat | Store Bat | List Bat | Update Bat |
|---------|--------------|-----------|----------|------------|

Instrument Setup: %R LAV SCI UV Cal

Date: 05-Aug-23 Time: 10:57:55

CMC DE Pass/Fail Tolerance: 1.00


I:c Ratio: (2.00 1.00)

Measured by:

| Batch Name | III/Obs | DL* | Da* | Db* | DL*/ISL | DC*/cSC | DH*/SH | CMC DE | M.I. | P/E | Color | Comments | |
|--|------------|-------|------|------|---------|---------|--------|--------|------|------|----------|------------|---------|
| DKF19106 230120130IR-01 83697 EXP23006050 MOJAVE T | D65 10 Deg | -1.93 | 0.41 | 0.40 | -0.75 | 0.37 | -0.50 | 0.97 | | Warn | too dark | too bright | too red |
| DKF19106 230120130IR-01 83697 EXP23006050 MOJAVE T | F11 10 Deg | -1.88 | 0.12 | 0.44 | -0.72 | 0.35 | -0.11 | 0.81 | 0.32 | Warn | too dark | too bright | too red |
| DKF19106 230120130IR-01 83697 EXP23006050 MOJAVE T | A 10 Deg | -1.86 | 0.31 | 0.57 | -0.72 | 0.50 | -0.18 | 0.89 | 0.21 | Warn | too dark | too bright | too red |

DataColor TOOLS 2.8.1

S/N NO. 8823242



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form