

Standard Name 110 of 140

↑

7091/210338D-C

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 14-Dec-22 Time: 18:03:30

Batch Name 2 of 2

↑

R2218265 5366657 92432 DOM2300842 JAVA

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 12-Sep-23 Time: 09:46:43


CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)

Measured by:

▼

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
R2218265 5366657 92432 DOM2300842 JAVA	D65 10 Deg	-1.39	0.15	0.03	-0.91	0.10	-0.12	0.93		Warn	too dark	too bright	too red
R2218265 5366657 92432 DOM2300842 JAVA	F11 10 Deg	-1.37	0.33	0.04	-0.89	0.18	-0.39	0.99	0.16	Warn	too dark	too bright	too red
R2218265 5366657 92432 DOM2300842 JAVA	A 10 Deg	-1.38	0.19	0.05	-0.87	0.12	-0.15	0.89	0.16	Warn	too dark	too bright	too red



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form