

Standard Name    390   of   395

↑

17080/220760M-C

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV   SCI UV Cal

Date: 11-Nov-22                      Time: 15:33:57

Batch Name                      1       of       2

↑

3088011AA-001 75389 EXP2300414 LIGHT QU

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV   SCI UV Cal


Date: 16-Jul-23                      Time: 14:21:22

CMC DE Pass/Fail Tolerance:       0.80

I:c Ratio: (   2.00   1.00 )

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
MTR17080 B23088011AA-001 75389 EXP2300414 LIGHT QU	D65 10 Deg	-0.43	-0.16	0.25	-0.18	-0.04	0.34	0.39		Pass	too dark	too dull	too yellow
MTR17080 B23088011AA-001 75389 EXP2300414 LIGHT QU	F11 10 Deg	-0.44	-0.20	0.23	-0.18	-0.06	0.36	0.41	0.04	Pass	too dark	too dull	too yellow
MTR17080 B23088011AA-001 75389 EXP2300414 LIGHT QU	A 10 Deg	-0.43	-0.19	0.24	-0.18	-0.03	0.36	0.41	0.02	Pass	too dark	too dull	too yellow



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form