

Standard Name14 of 182

↑

19106/230094M-A

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAVSCI UV Cal

Date: 19-May-23Time: 11:25:04

Batch Name5 of 5

↑

1020059IR-01 107635 EXP2300818 compass kh

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAVSCI UV Cal

Date: 23-Nov-23Time: 14:08:06

CMC DE Pass/Fail Tolerance:1.00

I:c Ratio: (2.001.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
SJS7091 24020059IR-01 107635 EXP2300818 compass kh	D65 10 Deg	-1.86	0.48	0.03	-0.77	0.04	-0.58	0.97		Warn	too dark	too bright	too red
SJS7091 24020059IR-01 107635 EXP2300818 compass kh	F11 10 Deg	-1.88	0.53	0.01	-0.77	0.02	-0.60	0.98	0.05	Pass	too dark	too bright	too red
SJS7091 24020059IR-01 107635 EXP2300818 compass kh	A 10 Deg	-1.81	0.47	0.11	-0.74	0.16	-0.56	0.94	0.15	Pass	too dark	too bright	too red

