

Standard Name60 of 60

ABZB True Pink-SS19

Visual Hue Request: RED/YELLOWISH

New Std	Retrieve Std	Store Std	List Std
---------	--------------	-----------	----------

Instrument Setup:%R LAV SCI UV Cal

Date: 03-Jan-22Time: 16:03:12

Batch Name1 of 1


7A/E2089 86131 DOM2300881 ABZB TRU

New Bat	Retrieve Bat	Store Bat	List Bat
---------	--------------	-----------	----------

Instrument Setup: %R LAV SCI UV Cal

Date: 18-Aug-23Time: 09:15:06

DeltaD65 / 10F11 / 10



Single/All	Zoom In	Zoom Out	CMC Plot	Print Form
------------	---------	----------	----------	------------

DataColor TOOLS 2.8.1S/N NO. 8823242

Standard %R

400-450nm	69.71	70.18	68.17	65.00	61.81	57.68
460-510nm	55.01	53.12	49.46	46.54	45.08	42.57
520-570nm	40.37	40.97	42.80	43.01	42.98	48.06
580-630nm	57.87	67.87	75.31	78.42	78.49	78.02
640-690nm	77.67	78.51	82.45	86.17	91.97	90.97
700nm	91.49					

CMC DE Pass/Fail Tolerance: 0.80

I:c Ratio: (2.00 1.00)

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments	
62733223 POAO2307A/E2089 86131 DOM2300881 ABZB TRU D65 10 Deg	-0.13	-0.39	0.77	-0.05	-0.29	0.50	0.58			Pass	too dark	too dull	too red
62733223 POAO2307A/E2089 86131 DOM2300881 ABZB TRU F11 10 Deg	-0.36	1.16	0.65	-0.13	0.58	0.61	0.85	1.58	Fail	too dark	too bright	too red	