



a xylem brand

# CALIBRATION CERTIFICATE

Form No 854, Dec 16th 2016

Certificate no: Oxygen Optode 4831\_476\_127805  
FoilID: 1206EM

Product: Oxygen Optode 4831  
Serial no: 476  
Calibration date: 01.03.2017

## 2-POINT RECALIBRATION, MULTIPOINT CALIBRATED OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

### Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 $\mu$ M	0 - 120%	-5 to +40°C
Accuracy	< $\pm 3\mu$ M or $\pm 2\%$ 1)	$\pm 2\%$ 2)	$\pm 0.03^\circ$ C 3)
Resolution	< 1 $\mu$ M	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter

2) Within calibrated range 0 - 120% / 0 - 30°C

3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

### Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	33.65	61.83
Temperature reading (°C)	10.01	21.83
Air Pressure (hPa)	989.00	

### Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-1.415263E+00	1.053662E+00

### With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.89149E-03	1.23681E-04	2.40123E-06	2.30565E+02	-3.40923E-01	-5.92159E+01	4.53423E+00
PhaseCoef	0.000000E+00	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.595317E+01	-3.090192E-02	3.027413E-06	-4.462716E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						

Date:01.03.2017

  
Sr. Repair Technician