



## SAMI<sup>2</sup> pH Validation Certificate

November 1, 2021

Serial Number: P131 RMA Number: 091340 Technician: BW

Reagent: pmcp0866

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.5

## Validation Notes

Tris Batch: 092921  $$\Delta\,\mathrm{T}=0.16\,^{\circ}\mathrm{C}$$ 

Tris Bottle: 6 N=9

Validation Temp: 24.97 °C  $\Delta pH = 0.0026 \pm 0.0008$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Standard refurbishment

OOI IR testing Filters replaced