



Sunburst Sensors, LLC
1226 W Broadway Ave
Missoula, MT 59802
+1 406 532 3247

SAMI² CO₂ Calibration Certificate

May 13, 2016

Serial Number: C64
Technician: RC Darlington
Reagent: 1/27/16

Table 1: SAMI² CO₂ calibration parameters.

A	B	C	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0655	0.5046	-1.4010	1.1138	0.9326	15.5494	201–1190

Table 2: Blank intensities.

Source	Reference	Signal
Dark	42	106
434 nm	2486	2830
620 nm	3060	2927

Build Notes

Pressure Test: 200 psi

Modifications:

