



Sunburst Sensors, LLC  
1226 W Broadway Ave  
Missoula, MT 59802  
+1 406 532 3246

## SAMI<sup>2</sup> pH Validation Certificate

November 20, 2023

Serial Number: P0133  
RMA Number: 280  
Technician: BW  
Reagent: pmcp1201

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 111023

$\Delta T = 0.08^{\circ}\text{C}$

Tris Bottle: 2

N = 11

Validation Temp: 25.03 °C

$\Delta \text{pH} = -0.0047 \pm 0.0008$

## Build Notes

Pressure Test: 200 psi – Passed

Modifications:

standard refurb

IR testing

2x sem u/g