



## SAMI<sup>2</sup> pH Validation Certificate

February 14, 2017

Serial Number:

P0154

Technician:

B. Wasser

Reagent:

pmp0129

Table 1:  $SAMI^2$  pH E-values.

$Ea_{434}$	$\mathrm{Eb}_{434}$	$\mathrm{Ea}_{578}$	$\mathrm{Eb}_{578}$
17372.0	2284.1	94.1	38676.5

## Validation Notes

Tris Batch: 111616

 $\Delta T = 0.06$  °C

Tris Bottle: 7

N = 9

Validation Temp: 24.86 °C

 $\Delta \, \mathrm{pH} = -0.0002 \pm 0.0006$ 

## **Build Notes**

Pressure Test: 200 psi-Passed

Modifications:

Swagelok ventable plug added

Firmware upgraded to v 55

Peek mixing tube added/replaced