



## SAMI<sup>2</sup> pH Validation Certificate

November 22, 2023

Serial Number: P0163 RMA Number: 280 Technician: BW

Reagent: pmcp1202

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 111023  $\Delta T = 0.07 \,^{\circ}\text{C}$ 

Tris Bottle: 3 N = 10

Validation Temp: 25.05 °C  $\Delta pH = -0.0046 \pm 0.0006$ 

## **Build Notes**

 $Pressure\ Test:\ 200\,psi-Passed$ 

 ${\bf Modifications:}$ 

standard refurb

IR testing

Replaced damaged board, 2x sem ug