



May 11, 2015

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0159

OOI Identification Number: PH4DO4

CGSN P/N: 1336-00013-000004

Date: 05/11/15

pH Evaluates:

$Ea_{434} = 17533$

$Ea_{578} = 101$

$Eb_{434} = 2229$

$Eb_{578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$T_{\text{offset}} = 0.07\text{ C}$

Certified Tris Batch #: Tris 042315

Tris Bottle/Bag #: Bottle

Tris Correction: 1

mCP lot: MKBC2604V

Date: 05/11/15

Pressure Test:

200psi – passed

Technician: J. Seidel

Modifications: 1-Liter Reagent bag