



SAMI^2 pH Validation Certificate

August 2, 2023

Serial Number: P0159 RMA Number: 082311 Technician: BP

Reagent: PmCP1137

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 061423 $\Delta T = 0.11\,^{\circ}\text{C}$

Tris Bottle: N = 9

Validation Temp: 24.98 °C $\Delta pH = 0.0018 \pm 0.0006$

Build Notes

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$

Standard Refurb

Replaced LED and filter