



## $\mathrm{SAMI}^2$ pH Validation Certificate

April 2, 2024

Serial Number: P0117 RMA Number: 281 Technician: BW

Reagent: pmcp1289

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 013124  $$\Delta\,\mathrm{T}=0.16\,^{\circ}\mathrm{C}$$ 

Tris Bottle: 9 N=9

Validation Temp: 24.98 °C  $\Delta pH = -0.0009 \pm 0.0006$ 

## **Build Notes**

 $Pressure\ Test:\ 200\,psi-Passed$ 

 ${\bf Modifications:}$ 

standard refurbishment

IR testing

2x Semrock filter upgrade