



## SAMI<sup>2</sup> pH Validation Certificate

November 23, 2020

Serial Number: P0110 RMA Number: 091756

> Technician: B. Montgomery Reagent: PmCP\_0732

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 091620  $\Delta T = 0.02 \,^{\circ}\text{C}$ 

Tris Bottle: 5 N=9

Validation Temp: 25.14 °C  $\Delta pH = -0.0013 \pm 0.0011$ 

## **Build Notes**

 $Pressure\ Test:\ 200\,psi-Passed$ 

 ${\bf Modifications:}$ 

Standard refurbishment

Insulation resistance testing