



SAMI^2 pH Validation Certificate

September 6, 2018

Serial Number: P0159 RMA Number: 071456 Technician: B. Wasser Reagent: pmcp0393

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 073118 $$\Delta\,\mathrm{T}=0.10\,^{\circ}\mathrm{C}$$

Tris Bottle: 5 N=9

Validation Temp: $25.22\,^{\circ}$ C $\Delta pH = 0.0034 \pm 0.0008$

Build Notes

 $Pressure\ Test:\ 200\,psi-Passed$

 ${\bf Modifications:}$

faulty valve replaced