

${\rm SAMI^2~CO_2~Calibration~Certificate}$

November 8, 2021

Serial Number: c80

RMA Number:

OOI Number: 091331 Technician: BW Reagent: btb0404

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	С	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0805	0.4572	-1.3710	0.9590	1.4833	12.6857	206-1957

Table 2: Blank intensities.

Source	Reference	Signal
Dark	72	63
$434~\mathrm{nm}$	2936	2816
$620~\mathrm{nm}$	1914	2803

Build Notes

Pressure Test: 200 psi

Modifications:

Standard Refurbishment

OOI IR testing

 $Pump\ replacement\ (low\ flow)$

434 LED, 620 filter repalced

extended calibration range

