



## $\mathrm{SAMI}^2$ pH Validation Certificate

October 11, 2019

Serial Number: P0112 RMA Number: 082720

> Technician: B. Montgomery Reagent: PmCP\_0566

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
18103.0	2296.0	97.8	40427.0

## Validation Notes

Tris Batch: 070319  $\Delta T = 0.14\,^{\circ}\text{C}$ 

Tris Bottle: 3 N=9

Validation Temp: 24.97 °C  $\Delta pH = -0.0041 \pm 0.0007$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Standard refurbishment