

## ${\rm SAMI^2~CO_2~Calibration~Certificate}$

December 4, 2023

Serial Number: SAMI C0047

RMA Number:

OOI Number: 276
Technician: BW

Reagent: btb\_0534

Table 1:  $SAMI^2 CO_2$  calibration parameters.

A	В	$^{\mathrm{C}}$	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.0719	0.4877	-1.3082	1.1370	0.9532	7.5356	202-1191

Table 2: Blank intensities.

Source	Reference	Signal
Dark	30	37
$434~\mathrm{nm}$	2603	2962
$620~\mathrm{nm}$	2857	2731

## **Build Notes**

Pressure Test: 200 psi

Modifications:

standard refurb

IR testing

ext cal

