



Sunburst Sensors, LLC  
1226 W Broadway Ave  
Missoula, MT 59802  
+1 406 532 3246

## SAMI<sup>2</sup> pH Validation Certificate

August 27, 2018

Serial Number: P0162  
RMA Number: 062916  
Technician: B. Wasser  
Reagent: pmcp0394

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.1	94.1	38676.5

## Validation Notes

Tris Batch: 073118

$\Delta T = 0.08^{\circ}\text{C}$

Tris Bottle: 3

N = 9

Validation Temp: 25.13 °C

$\Delta \text{pH} = 0.0022 \pm 0.0007$

## Build Notes

Pressure Test: 200 psi – Passed

Modifications:

standard service

60% reagent rem.