



## $\mathrm{SAMI}^2$ pH Validation Certificate

October 18, 2024

Serial Number: P0161 RMA Number: 784 Technician: BW Reagent: 1390

Table 1: SAMI<sup>2</sup> pH E-values.

| Ea <sub>434</sub> | Eb <sub>434</sub> | Ea <sub>578</sub> | Eb <sub>578</sub> |
|-------------------|-------------------|-------------------|-------------------|
| 19168.0           | 2546.0            | 111.0             | 41962.0           |

## Validation Notes

Tris Batch: 071524  $\Delta T = 0.12$  °C

Tris Bottle: 7 N=7

Validation Temp: 24.94 °C  $\Delta pH = -0.0004 \pm 0.0005$ 

## **Build Notes**

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$ 

standard refurb 2x Sem filters IR testing