



## SAMI<sup>2</sup> pH Validation Certificate

May 4, 2023

Serial Number: P0132 RMA Number: 083125 Technician: BW

Reagent: pmcp1086

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 041123  $\Delta T = 0.14 \,^{\circ}\text{C}$ 

Tris Bottle: 5 N=9

Validation Temp: 25.10 °C  $\Delta pH = 0.0030 \pm 0.0009$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Standard Refurbishment

IR testing New Pump

50 pct reagent