



Sunburst Sensors, LLC  
1226 W Broadway Ave  
Missoula, MT 59802  
+1 406 532 3247

## SAMI<sup>2</sup> CO<sub>2</sub> Calibration Certificate

August 30, 2018

Serial Number: c123  
RMA Number: 071404  
Technician: B. Wasser  
Reagent: btb0192

Table 1: SAMI<sup>2</sup> CO<sub>2</sub> calibration parameters.

A	B	C	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.0049	0.8800	-1.9655	0.9607	1.4255	15.5631	200–1196

Table 2: Blank intensities.

Source	Reference	Signal
Dark	83	104
434 nm	3091	2995
620 nm	2209	3135

## Build Notes

**Pressure Test:** 200 psi

**Modifications:**

standard service

extended cal

