



April 24, 2019

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0153

OOI Identification Number: PH4301

RSN P/N: 4830-58337

Date: 03/17/2015

pH Evaluates:

$Ea_{434} = 17533$

$Ea_{578} = 101$

$Eb_{434} = 2229$

$Eb_{578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$T_{\text{offset}} = 0.10\text{ C}$

Certified Tris Batch #: Tris 012715

Tris Bottle/Bag #: Bottle

Tris Correction: 0

mCP lot: MKBC2604V

Date: 03/17/2015

Pressure Test:

200psi – passed

Technician: J. Seidel

Modifications: Reagent Extension for 1.5-L bag