



SAMI^2 pH Validation Certificate

October 18, 2024

Serial Number: P0059 RMA Number: 784 Technician: BW Reagent: 1386

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
19168.0	2546.0	111.0	41962.0

Validation Notes

Tris Batch: 071524 $\Delta\,\mathrm{T} = 0.12\,\mathrm{^{\circ}C}$

Tris Bottle: 7 N=7

Validation Temp: 24.94 °C $\Delta pH = 0.0017 \pm 0.0007$

Build Notes

Pressure Test: 200 psi – Passed

Modifications:

standard refurb 2x sem filters

new diaphtagm

IR testing