



SAMI^2 pH Validation Certificate

November 20, 2023

Serial Number: P0115 RMA Number: 280 Technician: BW

Reagent: pmcp1200

Table 1: SAMI² pH E-values.

| Ea ₄₃₄ | Eb ₄₃₄ | Ea ₅₇₈ | Eb ₅₇₈ |
|-------------------|-------------------|-------------------|-------------------|
| 17372.0 | 2284.0 | 94.1 | 38676.0 |

Validation Notes

Tris Batch: 111023 $\Delta T = 0.14\,^{\circ}\text{C}$

Tris Bottle: 2 N=9

Validation Temp: 25.06 °C $\Delta pH = -0.0032 \pm 0.0013$

Build Notes

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$

standard refurm

IR testing

2x sem u/g