



SAMI² pH Validation Certificate

November 11, 2019

Serial Number: P131 RMA Number: 082727 Technician: BW

Reagent: pmcp0583

Table 1: $SAMI^2$ pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.5

Validation Notes

Tris Batch: 101519 $\Delta T = 0.12 \,^{\circ}\text{C}$

Tris Bottle: 4 N = 7

Validation Temp: 25.22 °C Δ pH = -0.0027 ± 0.0008

Build Notes

 $Pressure\ Test:\ 200\,psi-Passed$

 ${\bf Modifications:}$

refurbishment with OOI 3-pin upgrade; pv cable, bulkhead, pressure plate and labor.

Replacement of pump and valve

75% reagent remained on rcpt