

${\rm SAMI^2~CO_2~Calibration~Certificate}$

February 13, 2024

Serial Number: c80
RMA Number: 277
OOI Number: NA
Technician: BW
Reagent: btb0551

Table 1: $SAMI^2 CO_2$ calibration parameters.

| A | В | С | K_{434} | K_{620} | T_{avg} | Range (ppm) |
|--------|--------|---------|-----------|-----------|-----------|-------------|
| 0.0479 | 0.6399 | -1.7264 | 1.1823 | 1.7503 | 8.2363 | 208-1992 |

Table 2: Blank intensities.

| Source | Reference | Signal |
|-------------------|-----------|--------|
| Dark | 66 | 69 |
| $434~\mathrm{nm}$ | 2457 | 2895 |
| 620 nm | 1556 | 2680 |

Build Notes

Pressure Test: 200 psi

Modifications:

standard Refurb

IR testing

