



SAMI^2 pH Validation Certificate

November 1, 2021

Serial Number: P0164 RMA Number: 091340 Technician: BW

Reagent: pmcp0867

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 092921 $$\Delta\,\mathrm{T}=0.16\,^{\circ}\mathrm{C}$$

Tris Bottle: 6 N=9

Validation Temp: 24.97 °C $\Delta pH = 0.0038 \pm 0.0007$

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

Standard Refurbishment

OOI IR testing