



## $\mathrm{SAMI}^2$ pH Validation Certificate

July 7, 2023

Serial Number: P0059 RMA Number: 082305 Technician: BW

Reagent: pmcp1126

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 061423  $\Delta T = 0.12 \,^{\circ}\text{C}$ 

Tris Bottle: 2 N=9

Validation Temp: 25.11 °C  $\Delta pH = 0.0024 \pm 0.0011$ 

## **Build Notes**

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$ 

standard refurbishment

IR testing

10 pct reagent