

$SAMI^2 CO_2$ Calibration Certificate

June 14, 2017

Serial Number: C0076 RMA Number: 4191 Technician: B. Wasser Reagent: btbeg0013

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	С	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0862	0.4189	-1.2160	1.2983	1.2699	7.9599	207–1194

Table 2: Blank intensities.

Source	Reference	Signal
Dark	45	78
$434~\mathrm{nm}$	2301	3005
$620~\mathrm{nm}$	2327	2977

Build Notes

Pressure Test: 200 psi

Modifications:

Signals needed to be adjusted. Recalibration required.

Cold Cal, extended range 1200-200

3-pin P/V cable replaced per RMA

