



SAMI² pH Validation Certificate

October 27, 2021

Serial Number: P0116 RMA Number: 091340 Technician: BW

Reagent: pmcp0861

Table 1: SAMI² pH E-values.

| Ea ₄₃₄ | Eb ₄₃₄ | Ea ₅₇₈ | Eb ₅₇₈ |
|-------------------|-------------------|-------------------|-------------------|
| 17372.0 | 2284.0 | 94.1 | 38676.5 |

Validation Notes

Tris Batch: 092921 $\Delta T = 0.15 \,^{\circ}\text{C}$

Tris Bottle: 5 N=7

Validation Temp: 25.05 °C $\Delta pH = -0.0006 \pm 0.0010$

Build Notes

Pressure Test: 200 psi - Passed

Modifications:

Standard Refurbishment

Replaced Damaged swagelok (Corrosion) $\,$

Repaired damaged pressure cap (1hr)

OOI IR testing