

$SAMI^2 CO_2$ Calibration Certificate

October 16, 2017

Serial Number: C75
RMA Number: 7160
Technician: B. Wasser
Reagent: BTB0126

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	\mathbf{C}	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0687	0.4969	-1.3721	1.0535	1.1340	11.9303	100-1199

Table 2: Blank intensities.

Source	Reference	Signal
Dark	84	117
$434~\mathrm{nm}$	2749	2926
$620~\mathrm{nm}$	2480	2837

Build Notes

Pressure Test: 200 psi

Modifications:

Flooded SAMI rebuild extended range cal 1200uATM Extra lg CO2 reagent

