

## $SAMI^2 CO_2$ Calibration Certificate

May 8, 2019

Serial Number: c117

RMA Number:

Technician: B Wasser Reagent: BTB040919

Table 1:  $SAMI^2 CO_2$  calibration parameters.

A	В	$\mathbf{C}$	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.1107	0.2554	-1.0417	0.7089	0.9601	17.1328	201–1199

Table 2: Blank intensities.

Source	Reference	Signal
Dark	34	59
434 nm	3522	2530
620  nm	2757	2676

## **Build Notes**

Pressure Test: 200 psi

**Modifications:** 

Recalibration with new reagent-no refurbishment

