



May 7, 2014

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0133

OOI Identification Number:

RSN P/N: 4830-58337

RSN AT: ATAPL58337-00006

Date: 5/7/14

pH Evaluates:

$Ea_{434} = 17533$

$Ea_{578} = 101$

$Eb_{434} = 2229$

$Eb_{578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$T_{\text{offset}} = 0.15\text{C}$

Certified Tris Batch #: 042214

Tris Bottle/Bag #: bottle

Tris Correction: 0.0028

mCP lot: MKBC2604V

Date: 5/5/14

Pressure Test:

200psi – passed

Modifications:

Pump/valve bubble trap

Modified diaphragm retainer ring