



In Water Digital Linear PAR Sensor Calibration

Date: January 11, 2016
Customer: University of Washington
Project Number: 2015-3273
Model: PAR LIN 600m
Serial Number: 434

Calibration Coefficients

$I_m = 1.3589$
 $a_0 = 2156586820$ [count]
 $a_1 = 2.507589e-006$ [$\mu\text{mol photons/m}^2/\text{s/count}$]

Equation: $y = I_m * a_1 (x - a_0)$

where,
PAR is y [$\mu\text{mol photons/m}^2/\text{s}$]
Digital signal is x [count]

Table 1 - Connector Pin Descriptions

Pin	Signal	Description
1	GND	Power Supply Return
2	RX	RS232
3	TX	RS232
4	Power	Power
5	NC	Not Connected
6	NC	Not Connected



MCBH-6F
Female Face View

Satlantic
Richmond Terminal, Pier 9
3481 North Marginal Road
Halifax, Nova Scotia
B3K 5X8
Canada

Tel: 902-492-4780
Fax: 902-492-4781