

CALIBRATION CERTIFICATE

Form No 830, Juli 2012

a **xylem** brand

Certificate no: 4831_125_00077856 Foil batch no: 1206EM

Product: 4831

Calibration date: 19.12.2012

Serial no: 125 Page 1 of 2

Index	Temperature reference(°C)	[O2] Reference(µM)	Temperature raw data(mV)	Phase reading(°)
0	30.411	0.95	-134.011	60.71
1	20.153	0.20	192.099	61.63
2	9.827	0.07	524.024	62.40
3	0.603	0.38	800.570	63.03
4	0.804	19.47	794.889	60.39
5	0.905	41.34	792.031	57.66
6	0.973	61.78	790.124	55.38
7	1.024	103.29	788.655	51.37
8	1.059	145.15	787.675	48.00
9	1.093	219.02	786.695	43.27
10	1.110	322.73	786.216	38.40
11	1.120	429.76	785.918	34.74
12	1.127	529.84	785.729	32.12
13	10.678	14.29	497.229	59.51
14	10.606	31.81	499.518	. 56.43
15	10.542	48.29	501.553	53.88
16	10.494	82.22	503.059	49.45
17	, 10.466	113.93	503.970	46.08
18	10.448	175.16	504.504	41.02
19	10.439	257.50	504,818	36.18
20	10.429	341.25	505.112	. 32.66
21	10.435	414.46	504.918	30.32
22	20.384	11.76	184.596	58.47
23	20.320	25.00	186.674	55.27
24 .	20.270	39.14	188.303	52.32
25	20.233	65.34	189.492	47.77
26	20.156	94.63	191.973	43.78
27	20.138	140.82	192.567	39.00
28	20.128	205.73	192.893	34.28
29	20.122	272.72	193.073	30.87
30	20.107	336.54	193.563	28.45
31	30.370	9.71	-132.761	57.52
32	30.365	20.80	-132.604	53.96
33	30.366	31.67	-132.636	50.95
34	30.369	53.78	-132.750	45.99
35	30.381	75.74	-133.102	42.13
36	30.380	116.38	-133.074	36.94
37	30.379	171.22	-133.036	32.22
38	30.379	226.68	-133.041	28.92
39	30.378	278.22	-133.014	26.71



CALIBRATION CERTIFICATE

Form No 830, Juli 2012

a xylem brand

Certificate no: 4831_125_00077856


Foil batch no: 1206EM


Product: 4831

Calibration date: 19.12.2012

Serial no: 125 Page 2 of 2

Index	0	1	2	3	4	5	6
SUVFoilCoef	2.72962E-03	1.14611E-04	2.21263E-06	2.33768E02	-2.94251E-01	-5.33864E01	4.55129E00
TempCoef	2.61210E01	-3.15075E-02	3.14144E-06	-4.49045E-09	0.00000E00	0.00000E00	





Date:19.12.2012

Tor Duc Horlvog

Tor-Ove Kvalvaag, Calibration Engineer