



SAMI^2 pH Validation Certificate

October 20, 2021

Serial Number: P0115 RMA Number: 091340 Technician: BW

Reagent: pmcp0852

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: Dickson $\Delta T = 0.07 \,^{\circ}C$

Tris Bottle: T34 N=5

Validation Temp: 25.12 °C $\Delta pH = -0.0031 \pm 0.0016$

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

Standard Refurbishment

OOI IR Testing