



October 24, 2018

OOI SAMI-CO₂ Calibration Certificate

Sunburst SAMI Serial Number: C0080

OOI Identification #:

CGSN P/N: 1336-00012-000002

PCO2W Series B – PC2302

Date: 1/21/14

Calibration Range (ppm): 200-600

CO₂ Calibration Coefficients:

a = 0.0838

b = 0.2611

c = -0.8270

K₄₃₄ = 0.5609

K₆₂₀ = 1.0631

CO₂ Calibration Temperature

T = 15.79

Pressure Test:

200 psi

Modifications:

Pump/valve lid with bubble trap

Diaphragm retainer with ridge