



April 9, 2015

OOI SAMI-CO₂ Calibration Certificate

Sunburst SAMI Serial Number: C0116

OOI Identification #: RSN P/N: 4830-58336 / PC4301

Date: 04/09/15

Calibration Range (ppm): 200-600

CO₂ Calibration Coefficients:

$a = 0.0272$

$b = 0.6758$

$c = -1.6245$

$K_{434} = 0.7418$

$K_{620} = 1.2117$

CO₂ Calibration Temperature

$T = 14.6550$

Pressure Test:

200 psi