

${\rm SAMI^2~CO_2~Calibration~Certificate}$

December 4, 2023

Serial Number: SAMI C0091

RMA Number:

OOI Number: 276 Technician: BW

Reagent: btb_0531

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	С	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0696	0.4913	-1.3885	0.9546	1.1611	7.5670	269-1191

Table 2: Blank intensities.

Source	Reference	Signal
Dark	52	69
$434~\mathrm{nm}$	3016	2900
$620~\mathrm{nm}$	2577	3002

Build Notes

Pressure Test: 200 psi

Modifications:

standard refurb

 ${\rm IR} \ {\rm testing}$

Ext Cal

434 Sem u/g

