

${\rm SAMI^2~CO_2~Calibration~Certificate}$

December 30, 2019

Serial Number: c47

RMA Number: NA Technician: BW

Reagent: btb0271

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	C	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0792	0.4837	-1.3245	0.8713	0.8624	3.7160	202-1497

Table 2: Blank intensities.

Source	Reference	Signal 32	
Dark	29		
434 nm	2615	2282	
$620~\mathrm{nm}$	2496	2159	

Build Notes

Pressure Test: 200 psi

Modifications:

Re-Cal only 1500-200

