



July 5, 2012

## **SAMI-CO<sub>2</sub> Calibration Certificate**

**SAMI Serial Number: C0047**

**OOI/RSN identification number:**

**RSN P/N: 4830-58336**

**RSN AT: ATAPL58336-00001**

**Date: 7/2/2012**

**Technician: Reuben Darlington**

**Calibration Range (ppm): 200-600**

### **CO<sub>2</sub> Calibration Coefficients:**

$a = 0.0693$

$b = 0.4775$

$c = -1.3234$

$K_{434} = 0.9550$

$K_{620} = 1.1243$

### **CO<sub>2</sub> Calibration Temperature**

$T = 16.4$

$\text{Toffset} = 0$