



Sunburst Sensors, LLC
1226 W Broadway Ave
Missoula, MT 59802
+1 406 532 3246

SAMI² pH Validation Certificate

August 27, 2018

Serial Number: P0059
RMA Number: 071423
Technician: B. Wasser
Reagent: pmcp0397

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.1	94.1	38676.5

Validation Notes

Tris Batch: 073118

$\Delta T = 0.08^{\circ}\text{C}$

Tris Bottle: 3

N = 9

Validation Temp: 25.16 °C

$\Delta \text{pH} = 0.0029 \pm 0.0007$

Build Notes

Pressure Test: 200 psi – Passed

Modifications:

diaphragm damaged, valve corroded- replaced
still running on arrival 20% reagent